

PC850

Portable pH/Conductivity Meter

Manual

PH850 Portable pH Meter



EC850 Portable Cond. Meter



PC850 Portable pH/Cond. Meter



APERA INSTRUMENTS (Europe) GmbH

www.aperainst.de

Table of Contents

1. Introduction -----	1
2. Specifications -----	2
3. Instrument Description -----	3
3.1. LCD display-----	3
3.2. Keypad functions -----	3
3.3. Meter socket -----	5
3.4. Reading stability display mode -----	5
3.5. Automatic lock-up display mode -----	5
3.6. Automatic power off -----	5
3.7. Temperature adjustment -----	5
4. pH measurement -----	6
4.1. pH electrode information-----	6
4.2. pH calibration consideration -----	6
4.3. pH calibration -----	7
4.4. Sample test -----	8
4.5. pH electrode maintenance-----	9
5. mV value measurement-----	11
6. Conductivity measurement -----	12
6.1. Conductivity electrode information-----	12
6.2. Conductivity calibration -----	12
6.3. Conductivity calibration-----	13
6.4. Relationship between TDS and Conductivity -----	14
6.5. Sample test-----	14
6.6. Conductivity electrode maintenance-----	15
7. Parameter setting-----	15
7.1. Main menu-----	15
7.2. Submenu -----	16
7.3. Submenu of pH parameter setting -----	17
7.4. Submenu of Conductivity parameter setting -----	17
7.5. Submenu of standard basic parameter setting -----	18
8. Meter Kits-----	19
9. Warranty-----	19
Appendix I: Parameter setting & Factory default setting-----	20
Appendix II: Abbreviation Glossary-----	21
Appendix III: Self-diagnosis information-----	21

1. Introduction

Thanks for purchasing PC850 portable pH/Conductivity meter.

This meter is a perfect combination of the most advanced electronics technology, sensor technology, and software design, and is the most cost-effective portable electrochemical meter designed for industrial and mining enterprises, power plants, water treatment engineering, environmental protection industry, etc., especially ideal for general field use.

In order to use and maintain the instrument properly, please read the manual thoroughly before use. (To improve instrument performance constantly, we reserve the right to change the manual and accessories without giving notice in advance.)

1.1. Measurement Parameters

Measurement parameters	PH850	EC850	PC850
pH/mV	✓		✓
Conductivity		✓	✓
Temperature	✓	✓	✓

1.2. Basic Features

- The microprocessor-based portable meter features automatic calibration, automatic temperature compensation, function setup, self-diagnostics, automatic power-off and low voltage display. PC850 meter is able to measure up to 3 parameters simultaneously and switch displays effortlessly. Please see Diagram -1: two electrodes at most can be fit into the clip to measure 3 parameters simultaneously.
- The meter's digital filter improves measurement speed and accuracy. There is a stable reading indication on the display.
- The package includes a portable case, a meter, electrodes, standard solutions and all accessories, which is convenient to use in field.
- The meter is dust proof and waterproof, meeting the IP57 rating.



Diagram-1

1.3. pH Measurement Features (for Model PH850 and PC850)

- 1-3 points automatic calibration, the meter provides calibration guide and automatic checking function.
- The meter is able to recognize pH standard buffer solutions automatically. USA series and NIST series buffer solutions can be selected.
- The meter provides three types of reading stability criteria.

1.4. Conductivity Measurement Features (for Model EC850 and PC850)

- 1-4 points automatic calibration, the meter provides calibration guide and automatic checking function.
- The meter is able to switch between conductivity and TDS.
- The meter is able to recognize conductivity standard solutions automatically.

2. Specifications

2.1. Main specifications

	Specifications		Models
pH	Range	(0.00~14.00) pH	PH850 PC850
	Resolution	0.1/0.01 pH	
	Accuracy	±0.01 pH ±1 digit	
	Temperature compensation	(0~100)°C (manual or automatic)	
mV	Range	±1,000 mV	
	Resolution	1 mV	
	Accuracy	±0.2% F.S ±1 digit	
Conductivity	Range	Conductivity: 0~200 mS/cm (divided into five ranges): (0.00~19.99) µS/cm; (20.0~199.9) µS/cm; (200~1999) µS/cm; (2.00~19.99) mS/cm; (20.0~199.9) mS/cm TDS: (0~100) g/l	EC850 PC850
	Resolution	0.01/0.1/1µS/cm 0.01/0.1 mS/cm	
	Accuracy	±1.0% F.S ±1digit	
	Temperature compensation	(0 ~50)°C (manual or automatic)	
	Electrode constant	0.1 / 1 / 10 cm ⁻¹	
Temperature	Range	0~100°C	PH850 EC850 PC850
	Resolution	0.1°C	
	Accuracy	±0.5°C ±1 digit	

2.2. Other specifications:

Power	AAA batteries × 3 (1.5V×3)	
IP rating	IP57	
Dimension & Weight	Meter: (88×170×33) mm/313 g	
	Portable case:(360×270×76) mm/1.3 kg	PH850, EC850
	Portable case:(360×270×76) mm/1.4 kg	PC850

3. Instrument Description

3.1. LCD Display:




Diagram-2

- (1) — Parameter mode icons
- (2) — Measurement reading
- (3) — Prompts of special display mode
- (4) — Units of measurement
- (5) — Temperature units (°C and °F)
- (6) — Units of pH and conductivity calibration value
- (7) — pH and conductivity calibration value, and prompts of special display mode
- (8) — Temperature value, and prompts of special display mode
- (9) — Temperature compensation icons
 - ATC — automatic temperature compensation, MTC — manual temperature compensation
- (10) — Calibration guide icon
- (11) — Stability icon of readings
- (12) — Low battery icon, when this icon appears, please replace the battery
- (13) — Automatic reading hold icon

3.2. Keypad functions




Diagram-3

3.2.1. Keypad operations

Short press ----- <1.5 seconds; Long press ----- >1.5 seconds.

Turn on the meter: Press  to turn on the meter.

Turn off the meter: In the measurement mode, press  and hold for 2 seconds to turn off the meter.

Note: In the calibration mode or the parameter setup mode, pressing  is invalid. Please press  key to return to the measurement mode, then press  to turn off the meter.

Chart -1 Keypad operations and descriptions

Keypad	Operations	Descriptions
	Short press	● In the power-off mode, press this key to turn on the meter
	Long press	● In the measurement mode, press and hold this key for 2 seconds to turn off the meter.
	Short/long press	Select measurement parameters: ● pH 850 pH meter: pH → mV ● EC 850 Conductivity meter: COND → TDS ● PC 850 pH/Conductivity meter: pH → mV → COND → TDS ● In the measurement mode: long press to enter manual temperature compensation mode, then long press or momentary press this key to change the temperature value (only one direction). ● In the parameter setup mode, press this key to change the serial number of the main menu and the submenu (only one direction) ● In the submenu mode, press this key to change parameters and setup (only one direction)
	Short press	● In the measurement mode, press this key to enter in the calibration mode ● In the calibration mode or the parameter setup and auto lock-up (HOLD) mode, press this key to return to the measurement mode
	Short press	● In the measurement mode, press this key to enter in the parameter setup main menu ● In the calibration mode, press this key to make calibration ● In the parameter setup mode, press this key to select programs

3.3. Meter socket

Chart-2 Sockets for Meters

Models	Photos	Description
pH 850 pH meter		• BNC socket (right) — connect pH electrode or ORP electrode, • RCA socket (middle) — connect temperature probe
EC 850 Conductivity meter		• Eight-pin socket (left) — connect conductivity electrode
PC 850 pH/Conductivity meter		• BNC socket (right) — connect pH electrode or ORP electrode, • RCA socket (middle) — connect temperature probe • Eight-pin socket (left) — connect conductivity electrode

3.4. Reading Stability Display Mode

When the measuring value is stable, smiley icon ☺ appears on LCD, see Diagram – 4. If the smiley icon ☺ does not appear or flash, please do not get the reading value or make calibration until the measuring value is stable. Per parameter P1.3, there are 3 criteria for stability standard:

NOF (Normal), **HI** (High), and **Lo** (Low). The factory default is set “Normal”. “High” is set for stability for longer time, “Low” is set for stability for shorter time. User can select suitable stability criteria according to different testing requirement.




Diagram- 4

3.5. Automatic Lock-Up Display Mode

Select **On** from parameter P4.6 to turn on automatic lock-up display function. When the reading value stabilizes more than 10 seconds, the meter locks the measuring value automatically and displays **HOLD** icon, see Diagram – 5. In the **HOLD** mode, press  to release lock-up.




Diagram - 5

3.6. Automatic Power-off

The meter will be power-off if there is no operation for 20 minutes. The time of auto power-off can be set in parameter setting P3.2

3.7. Temperature Adjustment

When the temperature probe does not connect to the meter, long press  key, temperature value flashes, then long press  key or short press  key to change the temperature value in one way, and press  key to make confirmation.

4. pH Measurement

4.1. pH Electrode Information

The meter matches 201T-F plastic three-in-one combination pH electrode with built-in temperature sensor to realize automatic temperature compensation. Electrode housing adopts polycarbonate engineering plastics which is corrosion and impact resistant. The BNC socket of electrode connects pH socket; RCA socket connects temperature socket. When dip the electrode in the solution, please stir the solution briefly and allow it to stay in the solution until a stable reading is reached.

4.2. pH Calibration Consideration

4.2.1. Standard buffer solution

The meter uses two series of standard buffer solution: USA series and NIST series, 3 solutions for each series which are corresponding to calibration indication icons. Please see Chart - 3 for the two series of standard buffer solution.

Chart - 3 pH standard buffer solution series

Icons		pH standard buffer solution series	
		USA series	NIST series
Three-point calibration		4.00 pH	4.01 pH
		7.00 pH	6.86 pH
		10.01 pH	9.18 pH

4.2.2. Three-point calibration

The instrument can perform 1-3 points calibration. The first point calibration must use 7.00 pH (or 6.86 pH) standard solution, then select other standard solution to perform the second and the third point calibration. See chart – 4.

Chart - 4 Three-point calibration mode

	USA standard	NIST standard	Icons	Suited range
One-point calibration	7.00 pH	6.86 pH		Accuracy $\geq \pm 0.1\text{pH}$
Two-point calibration	7.00 pH $\rightarrow$ 4.00 pH	6.86 pH $\rightarrow$ 4.01 pH	 	Range < 7.00pH
	7.00 pH $\rightarrow$ 10.01 pH	6.86 pH $\rightarrow$ 9.18 pH	 	Range > 7.00pH
Three-point calibration	7.00 $\rightarrow$ 4.00 $\rightarrow$ 10.01 pH	6.86 $\rightarrow$ 4.01 $\rightarrow$ 9.18 pH	  	Large Range

4.2.3. Calibration intervals

Calibration intervals depend on the sample, the electrode performance, and the required accuracy. For high accuracy measurements ($\leq \pm 0.03\text{pH}$), the meter should be calibrated immediately before

taking a measurement. For general accuracy ($\geq \pm 0.1 \text{pH}$), the meter can be calibrated and used for approximately one week before the next calibration.

The meter must be recalibrated in the following situations:

- (a) New probe or probe that has not been used for a long time
- (b) After measuring acid ($\text{pH} < 2$) or alkaline solutions ($\text{pH} > 12$)
- (c) After measuring a solution that contains fluoride or a concentrated organic solution
- (d) If the solution's temperature differs greatly from the calibration solution temperature

4.3. pH Meter Calibration (with an example of three-point calibration)

4.3.1. First point calibration

(a) Press  key to enter into the calibration mode, “**CAL 1**” blinks at the top right of LCD and “7.00 pH” blinks at the bottom right of LCD, indicating using pH 7.00 buffer solution to make the 1st point calibration.

(b) Rinse pH electrode in pure water, allow it to dry, and submerge it in pH 7.00 buffer solution. Stir the solution briefly and allow it to stay in the buffer solution until a stable reading is reached. When stable  icon displays on LCD, press  key to calibrate, then 1st point calibration is finished, the meter enters in measurement mode of one-point calibration. Calibration guide icon  displays at the bottom left of LCD.

See Diagram-6.

4.3.2. Second point calibration

Press  key to enter into the calibration mode, “**CAL 2**” blinks at the top right of LCD indicating to make the 2nd point calibration. Rinse pH electrode in pure water, allow it to dry, and submerge it in pH 4.00 buffer solution. Stir the solution briefly and allow it to stay in the buffer solution until a stable reading is reached. The meter's display will show scanning and locking process of calibration buffer solution at the bottom right of LCD. When the meter locks 4.00 pH, stable  icon displays on LCD, press  key to calibrate the meter. LCD will display electrode slope of acidity range, then 2nd point calibration is finished, the meter enters in measurement mode of two-point calibration. Calibration guide icons  display at the bottom left of LCD.

4.3.3. Third point calibration

Press  key to enter into the calibration mode, “**CAL 3**” blinks at the top right of LCD indicating to make the 3rd point calibration. Rinse pH electrode in pure water, allow it to dry, and submerge it in




Diagram- 6

pH10.01 buffer solution. Stir the solution briefly and allow it to stay in the buffer solution until a stable reading is reached. The meter's display will show scanning and locking process of calibration buffer solution at the bottom right of LCD. When the meter locks 10.01pH, stable 😊 icon displays on LCD, press  key to calibrate the meter. LCD will display electrode slope of alkalinity range, then 3rd point calibration is finished, the meter enters in measurement mode of three-point calibration. Calibration guide icons  display at the bottom left of LCD.

Notes:

(a) The meter can perform 1-3 points calibration. When the 1st point calibration is done, the meter will enter in measurement mode of one-point calibration. Two-point and three-point calibration are in the same manner.

(b) During the calibration process, if measuring value is not stable yet (when 😊 does not appear and the  key is pressed, then **Er 2** displays. See chart – 5.

(c) To exit from the calibration mode, press  key.

4.4. Sample Test

4.4.1. Rinse pH electrode in pure water, allow it to dry, and submerge it in tested solution. Stir the solution briefly and allow it to stay in the tested solution until 😊 icon appears on LCD and a stable reading is reached, which is the pH value of tested solution. Diagram – 7 is the calibration and measurement process of pH meter







Diagram – 7 Calibration and measurement process of pH meter

4.4.2. Self-diagnosis information

During the process of calibration and measurement, the meter has self-diagnosis functions, indicating the relative information as below, please refer to chart – 5.

Chart – 5 Self-diagnosis information of pH measurement mode

Display Icons	Contents	Check and Fix
	Wrong pH buffer solution or the recognition of calibration solution out of range	1. Check whether pH buffer solution is correct. 2. Check whether the meter connects the electrode well. 3. Check whether the electrode is damaged.
	Press  key when measuring value is not stable during calibration.	Press  key after  icon appears
	During calibration, the measuring value is not stable for ≥3min.	1. Check whether there are bubbles in glass bulb. 2. Replace with new pH electrode.
	pH electrode performance error zero potential <-60mV or >60mV, slope <75%	1. Check whether there are bubbles in glass bulb. 2. Check whether pH buffer solution is correct. 3. Replace with new pH electrode.

4.4.3. pH temperature principle

Please note that the closer the temperature of the sample solution to the calibration solution, the more accurate the readings.

4.4.4. Factory default setting

For factory default setting, please refer to parameter setting P1.4 (Item 7.3). All calibration data is deleted and the meter restores to the theory value (zero electric potential of pH is 7.00, the slope is 100%). Some functions restore to the original value (refer to appendix -1). When calibration or measurement fails, please restore the meter to factory default setting and then perform re-calibration or measurement. Please note once set the factory default, all the data deleted will be irretrievable.

4.5. pH Electrode Maintenance

4.5.1. Daily maintenance

The soaking solution contained in the supplied protective bottle is used to maintain activation in the glass bulb and junction. Loosen the capsule, remove the electrode and rinse in pure water before taking a measurement. Insert the electrode and tighten the capsule after measurements to prevent the solution from leaking. If the soaking solution is turbid or moldy, replace the solution. (Preparation: 26g KCL dissolved in 100ml pure water).

The electrode should not be soaked in pure water, protein solution or acid fluoride solution for long periods of time. In addition, do not soak the electrode in organic silicon lipids.

For best accuracy, always keep the meter clean and dry, especially the and electrode connectors. Clean with medical cottons and alcohol if necessary.

4.5.2. Calibration buffer solution

For calibration accuracy, the pH of the standard buffer solution must be reliable. The buffer solution should be refreshed often, especially after heavy use.

4.5.3. Protect glass bulb

The sensitive glass bulb at the front of the combination electrode should not come in contact with hard surfaces. Scratches or cracks on the electrode will cause inaccurate readings. Before and after each measurement, the electrode should be washed with pure water and dried. Do not clean the glass bulb with a tissue for it will affect the stability of the electrode potential and increase the response time. The electrode should be thoroughly cleaned if a sample sticks to the electrode. Use a solvent if the solution does not appear clean after washing.

4.5.4. Renew glass bulb

Electrodes that have been used over a long period of time will become aged. Submerge the electrode in 0.1mol/L hydrochloric acid for 24 hours, then wash the electrode in pure water, then submerge it in soaking solution for 24 hours.

The method to prepare 0.1mol/L hydrochloric acid: dilute 9mL hydrochloric acid in pure water to 1000mL. For serious passivation, submerge the bulb in 4% HF (hydrofluoric acid) for 3-5 seconds, and wash it in pure water, then submerge it in the soaking solution for 24 hours to renew it.

4.5.5. Clean contaminated glass bulb and junction (please refer to Chart-6)

Chart – 6 Clean contaminated glass bulb and junction

Contamination	Cleaning Solutions
Inorganic metal oxide	Dilute acid less than 1mol/L
Organic lipid	Dilute detergent (weak alkaline)
Resin macromolecule	Dilute alcohol, acetone, ether
Proteinic haematocyte sediment	Acidic enzymatic solution (saccharated yeast tablets)
Paints	Dilute bleach, peroxide

Note: The electrode housing is polycarbonate. When use cleaning solutions, take cautions on carbon tetrachloride, trichlorethylene, tetrahydrofuran, acetone, etc., which will dissolve the housing and invalidate the electrode.

5. mV Value Measurement

5.1. Press  key, and switch the meter to mV measurement mode. Connect ORP electrode (need to purchase separately) and dip it in sample solution, stir the solution briefly and allow it to stay in the solution until  icon appears and get the reading, which is the ORP value.

ORP stands for Oxidation Reduction Potential. The unit is mV.

5.2. Notes

5.2.1. ORP measurement does not require calibration. When the user is not sure about ORP electrode quality or measuring value, use ORP standard solution to test mV value and see whether ORP electrode or meter works properly.

5.2.2. Clean and Activate ORP Electrode

After the electrode has been used over long period of time, the platinum surface will get polluted, which causes inaccurate measurement and slow response. Please refer to the following methods to clean and activate ORP electrode:

(a) For inorganic pollutant, submerge the electrode in 0.1mol/L dilute hydrochloric acid for 30 minutes, then wash it in pure water, then submerge it in the soaking solution for 6 hours.

(b) For organic or lipid pollutant, clean the platinum surface with detergent, then wash it in pure water, then submerge it in the soaking solution for 6 hours.

(c) For heavily polluted platinum surface on which there is oxidation film, polish the platinum surface with toothpaste, then wash it in pure water, then submerge it in the soaking solution for 6 hours.

6. Conductivity Measurement

6.1. Conductivity Electrode Information

6.1.1. Conductivity electrode

Model 2301T-S plastic conductivity electrode with constant $K=1.0$ and built-in temperature sensor, can realize automatic temperature compensation. The electrode housing is POM plastic which is corrosion resistant and impact resistant. When submerge the conductivity electrode in solution, stir the solution briefly to eliminate the air bubbles and improve response and stability.

6.1.2. Conductivity electrode constant

The meter matches conductivity electrodes of three constants: $K=0.1$, $K=1.0$ and $K=10.0$. Please refer to chart-7 for measuring range. Set constant in parameter setting P2.1 and refer to clause 7.4

Chart – 7 Electrode constant and measuring range

Range	< 20 $\mu\text{S/cm}$	0.5 $\mu\text{S/cm}$ ~100 mS/cm			> 100mS/cm
Conductivity electrode constant	$K=0.1 \text{ cm}^{-1}$	$K=1.0 \text{ cm}^{-1}$			$K=10 \text{ cm}^{-1}$
Standard solution	84 $\mu\text{S/cm}$	84 $\mu\text{S/cm}$	1413 $\mu\text{S/cm}$	12.88 mS/cm	111.9 mS/cm

6.2. Conductivity Calibration

6.2.1. Conductivity calibration solutions

The meter uses conductivity standard solution of 84 $\mu\text{S/cm}$, 1413 $\mu\text{S/cm}$, 12.88 mS/cm and 111.9 mS/cm. The meter can recognize the standard solution automatically, can perform one-point or multi-point calibration (the maximum is four-point calibration). The calibration indication icons correspond to the four standard values ( icon corresponds to two standards). See chart – 8:

Chart – 8 Conductivity standard solution series

Calibration guide icons	Calibration solution series	Range
	84 $\mu\text{S/cm}$	0-200 $\mu\text{S/cm}$
	1413 $\mu\text{S/cm}$	200-2,000 $\mu\text{S/cm}$
	12.88 mS/cm	2-20 mS/cm
	111.9 mS/cm	20-200 mS/cm

6.2.2. Calibration intervals

- (a) The meter is calibrated before leaving the factory and can generally be used right out of the box.
- (b) Normally perform calibration per month.
- (c) For high accuracy measurements or large temperature deviation from the reference temperature (25°C), perform calibration per week.
- (d) Use conductivity standard solution to check whether there is error. Perform calibration for large

error.

(e) For new electrode or factory default setting, perform 3-point or 4-point calibration. Choose closer standard solution to the sample solution to perform 1- point or 2-point calibration.

6.2.3. Reference temperature

Reference temperature of factory default is 25°C. Other reference temperature can also be set for range 15°C – 30°C. Select per parameter setting P2.2 and see clause 7.4.

6.2.4. Temperature coefficient

The temperature compensation coefficient of the meter setting is 2.0%/°C. However, the conductivity temperature coefficient is different from solutions and concentration. Please refer to chart – 9 and the data collected during testing. Set per parameter setting P2.3. and see clause 7.4.

Note: When the coefficient for the temperature compensation is set to 0.00 (no compensation), the measurment value will be based on the current temperature.

Chart -9 Temperature compensation coefficient of special solutions

Solution	Temperature compensation coefficient
NaCl solution	2.12%/°C
5% NaOH solution	1.72%/°C
Dilute ammonia solution	1.88%/°C
10% hydrochloric acid solution	1.32%/°C
5% sulfuric acid solution	0.96%/°C

6.2.5. Avoid contamination of standard solution

Conductivity standard solution has no buffer. Please avoid being contaminated during usage. Before submerging the electrode in standard solution, please wash the electrode and allow it dry. Please do not use the same cup of conductivity standard solution repeatedly, especially for standard solution of low concentration (84µS/cm). The contaminated standard solution will affect accuracy.

6.3. Conductivity Calibration (take 1413µS/cm as an example)

6.3.1. Rinse the electrode in pure water, allow it to dry, wash with a little of standard solution and submerge it in standard solution. Stir the solution briefly and allow it to stay in the solution until a stable reading is reached.

6.3.2. Press  key to enter into the calibration mode.

The meter's display will show blinking "CAL" at the top right, and scanning and locking process of calibration solution at the bottom right. When the meter locks 1413 mS,

stable icon ☺ will display on LCD. Press  key to complete calibration. The meter will return to measuring mode and  is displayed on bottom left of the LCD screen. See diagram-8.

6.3.3. For multi-point calibration, please repeat clause 6.3.1-6.3.2 until all the calibration is done. The meter can repeat calibration in the same calibration solution until the stable value is reached.

6.3.4. Notes:

- a) If press  key before stable icon ☺ appeared on LCD screen, Error indication icon Er 2 will be shown. See Chart -11.
- b) Press  key to exit calibration mode.

6.4. Relations between TDS and Conductivity

6.4.1. TDS and conductivity is linear related. The conversion factor is 0.40-1.00. Adjust the factor from parameter P2.4. The factory default setting is 0.71 and please refer to Item 7.4. The meter only needs to be calibrated in Conductivity mode, then after calibration of conductivity, the meter can switch from conductivity to TDS or salinity.

6.4.2 Adjust TDS conversion factor in parameter setting P2.4 according to the data collected during testing and experience. Chart – 10 lists some common use Conductivity and TDS conversion factors. This is for your reference only.

Chart – 10 Conductivity and TDS conversion factors

Conductivity of solution	TDS conversion factor
0-100 $\mu\text{S}/\text{cm}$	0.60
100-1,000 $\mu\text{S}/\text{cm}$	0.71
1-10 mS/cm	0.81
10-100 mS/cm	0.94

6.5. Sample test

6.5.1. Rinse conductivity electrode in pure water, allow it to dry, and submerge it in the sample solution. Stir the solution briefly and allow it to stay in the sample solution until a stable reading is reached and ☺ icon appears on LCD, then get the reading value, which is the conductivity value of the solution.



6.5.2. Press  key to switch to TDS.

6.5.3. During the process of calibration and measurement, the meter has self-diagnosis functions, indicating the relative information as below: chart – 11.



Diagram- 8

Chart – 11 Self-diagnosis information of conductivity measurement mode

Display Icons	Contents	Checking
	Wrong conductivity calibration solution or the meter recognition of calibration solution out of range	1. Check whether conductivity calibration solution is correct. 2. Check whether the meter connects the electrode well. 3. Check whether the electrode is damaged.
	Press  key when measuring value is not stable during calibration.	Press  key after  icon appears
	During calibration, the measuring value is not stable for ≥ 3 min.	1. Shake the electrode to eliminate bubbles in electrode head. 2. Replace with new electrode.

6.5.4 Factory default setting

For factory default setting, please refer to parameter setting P2.5 (Item 7.4). With this function, all calibration data is deleted and the meter restores to the theory value. Some functions restore to the original value (refer to appendix -1). When calibration or measurement fails, please restore the meter to factory default setting and then perform re-calibration or measurement. Please note once set the factory default, all the data deleted will be irretrievable.

6.6. Conductivity electrode maintenance

6.6.1. Always keep the conductivity electrode clean. Before taking a measurement, rinse the electrode in pure water and better to rinse it again in the sample solution. When submerge the electrode in solution, stir the solution briefly to eliminate air bubbles and allow it to stay until a stable reading is reached. For conductivity electrode which is drily stored, soak the electrode in pure water for 5-10 minutes. Rinse the electrode in pure water after measurement.

6.6.2. The interaction rod of Model 2301T-S conductivity electrode is coated with platinum black to minimize electrode polarization and expand measuring range. The platinum black coating adopted our special processing technology which improves the electrode performance and the firmness of the coating. If the platinum black electrode is stained, gently clean the electrode with soft brush in warm water containing detergent or alcohol.

7. Parameter setting

7.1. Main menu

In the measurement mode, press  key to enter in P1.0, then press  to switch to main menu:

P1.0→P2.0→P3.0. Please refer to diagram – 9.

P1.0: pH parameter setting menu,

P2.0: Conductivity parameter setting menu,

P3.0: Basic parameter setting menu.

7.2. Submenu

7.2.1. In P1.0 mode, press  key to enter in submenu P1.1 of pH parameter setting, then press  key to switch among submenu: P1.1→P1.2→P1.3→P1.4, see Diagram – 9.

7.2.2. In P2.0 mode, press  key to enter in submenu P2.1 of conductivity parameter setting, then press  key to switch among submenu: P2.1→P2.2→P2.3→P2.4→P2.5, see Diagram– 9.

7.2.3. In P3.0 mode, press  key to enter in submenu P3.1 of basic parameter setting, then press  key to switch among submenu: P3.1→P3.2→P3.3, see Diagram – 9.






Diagram – 9 Main menu and submenu of parameter setting

7.3. Submenu of pH parameter setting (press key to switch)



	P1.1. – Select pH standard solution (USA-NIST) 1. In measurement mode, press  key to enter in P1.0 mode, then press  to enter in P1.1. 2. When press  key, USA blinks, press  key to select USA→nIS, press  to confirm. USA-USA series; nIS-NIST series. 3. After confirm parameter, press  key to enter in P1.2 mode, or press  key to return to the measurement mode.
	P1.2. – Select resolution (0.01 – 0.1) 1. Press  key, 0.01 blinks, press  key to select 0.01→0.1, press  key to confirm. 2. After confirm parameter, press  key to enter in P1.3 mode, or press  key to return to the measurement mode.
	P1.3. – Set reading stability criteria (Normal – High – Low) 1. Press  key, nor blinks. Press  key to select nor→HI→Lo, press  to confirm. Nor – Normal, Hi – High, Lo – Low. 2. After confirm parameter, press  key to enter in P1.4 mode, or press  key to return to the measurement mode.
	P1.4. –Restore factory setting (No – Yes) 1. Press  key, no blinks. Press  key to select no→YES, press  to confirm, the meter returns to the measurement mode. No – Do not restore, Yes – Restore to factory setting. 2. If don't select Yes, press  key to return to the measurement mode.

7.4. Submenu of conductivity parameter setting (press key to switch)

	P2.1. – Select electrode constant (1.0-10.0-0.1) 1. In P2.0 mode, press  key to enter in P2.1 mode, please refer to the left Diagram. 2. Press  key, 1.0 blinks, then press  key to select 1.0→10.0→0.1, press  key to confirm. 3. After confirm the parameter, press  key to enter in P2.2 mode, or press  key to return to the measurement mode.
---	---

	P2.2. – Select reference temperature (15.0°C-30.0°C) 1. Press  key, 25.0°C blinks, then press  key to adjust temperature value 15.0-30.0, press  key to confirm. 2. After confirm parameter, press  key to enter in P2.3 mode, or press  key to return to the measurement mode.
	P2.3. – Adjust temperature compensation coefficient (0.00-9.99%) 1. Press  key, 2.00 blinks, press  key to adjust temperature compensation coefficient 0.00 – 9.99, press  key to confirm. 2. After confirm the parameter, press  key to enter in mode P2.4 or press  key to return to the measurement mode.
	P2.4. – Adjust TDS factor 1. Press  key, 0.71 blinks, press  key to adjust TDS factor 0.04 – 1.00, press  key to confirm. 2. After confirm the parameter, press  key to enter in mode P2.5 or press  key to return to the measurement mode.
	P2.5. –Restore factory setting (No – Yes) 1. Press  key, no blinks. Press  key to select no→YES, press  to confirm, the meter returns to the measurement mode. No – Do not restore, Yes – Restore to factory setting. 2. If don't select Yes, press  key to return to the measurement mode.

7.5. Submenu of basic parameter setting (press key to switch)

	P3.1. Select temperature unit (°C—°F). 1. In P3.0 mode, press  key to enter in P3.1 mode, please refer to the left Diagram. 2. Press  key, °C blinks, then press  key to select °C→°F, press  key to confirm. 3. When parameter is confirmed, press  key to enter in mode P3.2 or press  key to return to the measurement mode.
	P3.2 – Automatic power-off setup (10→20→30→On) 1. Press  key, On blinks, press  key to select 10→20→30→On Press  key to confirm. On – turn off automatic power-off; Time unit is minutes. 2. After confirm the parameter, press  key to enter in mode P3.3 or press  key to return to the measurement mode.

	P3.2 – Automatic lock-up setup (Off→On) 1. Press  key, OFF blinks, press  key to select OFF→On, press  to confirm. Off – not set; On–set (the reading is automatically locked when stabilizes > 10 seconds.) 2. When Parameter is confirmed, press  key to return to the measurement mode.
---	---

8. Meter Kits

No.	Include	Quantity	PH850	EC850	PC850
1.	PH 850 portable pH meter	1 set	√		
2.	EC 850 portable conductivity meter	1 set		√	
3.	PC 850 portable pH/conductivity meter	1 set			√
4.	201T-F plastic three-in-on pH electrode	1 pc	√		√
5.	2301T-S plastic conductivity electrode	1 pc		√	√
6.	pH standard buffer solution (4.00 pH /7.00pH/50mL)	1 bottle each	√		√
7.	Conductivity calibration solution (1413 µS/12.88 mS/50mL)	1 bottle each		√	√
8.	Combined electrode clip	1 pc			√
9.	Portable case	1 pc	√	√	√
10.	Manual	1 book	√	√	√

9. Warranty

We warrant this instrument to be free from defects in material and workmanship and agree to repair or replace free of charge any malfunctioned or damaged product attributable to the responsibility of APERA INSTRUMENTS for a period of THREE YEARS for the instrument and SIX MONTHS for the probe from the delivery.

This limited warranty does NOT cover any issues due to:

- Accidental damage, Improper use
- Normal wear and tear
- Transportation, Storage
- Failure to follow the product instructions
- Unauthorized maintenance, modifications, combination or use with any products, materials or other matter, Unauthorized repair
- External causes such as accidents, abuse, or other actions or events beyond our reasonable control.

Appendix I: Parameter setting & Factory default setting

Modes	Prompts	Parameter setting items	Abbreviation	Description	Restore to factory default
P1.0 pH	P1.1	Select pH buffer solution	buf	USA - NIST	USA
	P1.2	Select resolution	RES	0.01 - 0.1	0.01
	P1.3	Set reading stability criteria	SC	Normal—High—Low	Normal
	P1.4	Restore to factory default setting	FS	No - Yes	No
P2.0 Cond.	P2.1	Select electrode constant	CELL	1.0 - 10.0 - 0.1	1.0
	P2.2	Select reference temperature	REF	15~30°C	25°C
	P2.3	Adjust temperature compensation coefficient	TC	0.00~9.99	2.00
	P2.4	Adjust TDS factor	TDS	0.40~1.00	0.71
	P2.5	Restore to factory default setting	FS	No - Yes	No
P3.0 Basic Parameters	P3.1	Select temperature unit	/	°C - °F	°C
	P3.2	Automatic Power-off setup	AC	10 - 20 - 30 - On	20
	P3.3	Automatic Lock-up setup	/	/	Off

Appendix II: Abbreviation Glossary

Modes	Prompts	Code and abbreviation	In English	Description
P1.0 pH	P1.1	buf	Standard buffers	Standard buffer solution
	P1.2	rES	Resolution	Resolution
	P1.3	SC	Stability criteria	Set up reading stability criteria
	P1.4	FS	Factory default setting	Factory default setting
P2.0 Conductivity	P2.1	CELL	Cell	Constant Cell
	P2.2	ErEF	Reference temperature	Reference temperature
	P2.3	EC	Temperature compensation coefficient	Temperature compensation coefficient
	P2.4	TDS	Total dissolved solid	TDS
	P2.5	FS	Factory default setting	Factory default setting
P3.0 Basic parameters	P3.1	/	/	/
	P3.2	AC	Auto close	Automatic Power-off
	P3.3	/	/	/

Appendix III: Self-diagnosis information

Icons	Self-diagnosis information	pH	Conductivity
Er 1	Wrong pH buffer solution or the meter recognition of calibration solution out of range	√	√
Er 2	Press  key when measuring value is not stable during calibration	√	√
Er 3	During calibration, the measuring value is not stable for ≥ 3min.	√	√
Er 4	pH electrode performance error (zero potentia < -60mV or > 60mV, slope < 75%)	√	

APERA INSTRUMENTS (Europe) GmbH

Address: Wilhelm-Muthmann-Str.18

42329 Wuppertal Germany

Email: info@aperainst.de

Website: www.aperainst.de

Tel. +49 202 51988998